

Solicitation Amendment No. 001

Page 1 of 6

To:	Date:
Prospective Bidders:	October 22, 2008
Project Title:	Project No.:
Electric Hospital Beds	09-08
<p>Description of Solicitation Amendment: The Invitation for Bids (IFB) for Electric Hospital Beds (#09-08) is hereby amended as follows:</p> <p>I) Schedule of Items and Prices, Attachment No. 2:</p> <p>The Schedule of Items and Prices contained in the IFB is hereby deleted in its entirety and is replaced with the Revised Schedule of Items and Prices dated October 22, 2008, attached hereto and made a part hereof.</p> <p>II) The Bid Opening Date/Time is hereby extended until October 30, 2008 @ 2:00 P.M., local time.</p>	
Acknowledgement of Amendment No. 001 by:	Date:
Company Name (Bidder):	
Signed by:	
Name (Type or Print):	Title:

ATTACHMENT NO. 2
(Revised: October 22, 2008)

**SCHEDULE OF ITEMS AND PRICES
FOR
ELECTRIC HOSPITAL BEDS**

The Bidder/Contractor shall furnish all resources and services necessary and required to provide Electric Hospital Beds to Houston Community College in accordance with the general terms and conditions of the sample contract documents for the bid price(s) listed below. All equipment/beds shall be of new manufacture and shall not be used or refurbished.

<u>Item No.</u>	<u>Description</u>	<u>Qty</u>	<u>Unit</u>	<u>Unit Bid Price</u>
01	Electric Hospital Bed, w/standard mattress (CareAssist-ES Bed), Hill-Rom ES200 Package, or equal.	34	EA	\$_____
Bidding on: _____ (Manufacture/Model No.)				
02	Electric Hospital Bed, w/standard mattress (CareAssist-ES Bed), Hill-Rom ES220 Package, or equal.	2	EA	\$_____
Bidding on: _____ (Manufacture/Model No.)				
03	Electric Hospital Bed, w/standard mattress (VersaCare Bed), Hill-Rom Package VC220, or equal.	2	EA	\$_____
Bidding on: _____ (Manufacture/Model No.)				
04	Electric Hospital Bed, w/standard mattress (Bari 10A, Bariatric), Stryker Model Number 2230, Or Equal.	1	EA	\$_____
Bidding on: _____ (Manufacture/Model No.)				



05 Electric Hospital Bed, w/standard mattress 1 EA \$ _____
(Birthing), LD304, Stryker
Model Number 4701 or equal.

Bidding on: _____
(Manufacture/Model No.)

Training Services:

06 On-Site training services for faculty on 1 Day \$ _____
operation of beds identified in all line (8 Hours)
items listed above.

Total Bid Amount (Items 001 – 006): \$ _____

MINIMUM TECHNICAL SPECIFICATIONS
Line Item Nos. 1-5 of Attachment No. 2

Item #1 CareAssist ES200 or equal

Minimum Technical Specifications:

Frame width 39.74"
Frame length with head/ footboard (w/ bumpers)
normal 90.04"
extended 94.04"
Sleep deck
width 36.00"
normal length. 80.00"
extended length 84.00"
Bed clearance under frame. 5.00"
Low position floor to deck 15.75"
High position floor to deck 32.5"
Max head elevation. 65 degrees
Max Trend and reverse Trend 16 degrees
Caster size. 6"
Weight limit (Safe working load) 500 lbs

Item #2 CareAssist ES220 or Equal

Minimum Technical Specifications:

Frame width 39.74"
Frame length with head/ footboard (w/ bumpers)
normal 90.04"
extended 94.04"
Sleep deck
width 36.00"
normal length. 80.00"
extended length 84.00"
Bed clearance under frame. 5.00"
Scale System
accuracy +/- 2.2 lbs or 1.1% of patient weight w/e is greater
weigh in any position yes
scale repeatability 2% of patient weight
Low position floor to deck 32.5"
Max head elevation. 65 degrees
Max Trend and reverse Trend 16 degrees
Caster size. 6" (15 cm)
Weight limit (Safe working load) 500 lbs.

Item #3 **VersaCare VC220 or equal**

Minimum Technical Specifications:

Bed height

Low position – surface edge to floor.....18"
 with optional power transport22" (55.9 cm)
 High position – surface center to floor.....37"
 Headboard/push handles maintain same height regardless of patient surface height.....Yes

Sleep surface

Surface thickness.....8"
 Width.....35.5"
 Length, fully extended.....86"
 Length, fully retracted.....75"
 Non-radiolucent sleep deck.....Yes

Overall width

Siderails stowed.....37"
 Siderails up.....40"

Overall length

Fully extended.....97"
 Fully retracted.....85"

Bed angles

Head section.....0°-65°
 Knee gatch.....0°-16°
 Foot section.....0°-27°
 Max Trend/reverse Trend.....15° / 10°

Caster & braking system

Caster diameter.....5"
 with power transport.....6"
 Four-wheel braking.....Yes
 Brakes lock both rolling and swiveling.....Yes

Bed exit alarm system

Sensitivity levels.....3
 Modes.....Positioning, Exiting, Out-of-Bed

Scale system (standard)

Accuracy.....1% of patient weight
 Weigh in any position
 Scale weight capacity.....500 lbs.

Patient controls

Siderail patient bed controls..... Standard

Electrical characteristics

Backup battery power..... Standard

Item #4 Bari 10A, Bariatric, Stryker or equal

Minimum Technical Specifications:

- **Model Number:** 2230 – No Surface, 2231 – Foam, 2232 – Air
- **Overall Length:** Minimum: 90", Maximum: 98"
- **Overall Width:** Minimum: 39", Maximum: 52"
- **Weight Capacity:** 1000 lbs.
- **Bed Frame Weight:** 700 lbs
- **Standard Height Range**
High 33.7" / Low 18.7"
- **Litter Positioning**
Maximum back angle to horizontal 65°
Maximum knee angle to horizontal 35°
Maximum knee to leg angle 45°
Trend 14.5°
Reverse Trend 15.3°
- **Patient Surface**
Minimum 36" x 80" / Maximum 48" x 88"
- **Electronics**
Volts 110v AC
Ampere Rating 6 amps (115v)
Mains Voltage 110v AC
Frequency 60 Hz
Circuit Protection Auto reset current

Item #5: Birthing Bed, LD304, Stryker or equal

Minimum Technical Specifications

- **Model Number 4701 (International 4701 model has 230 volt option)**
- **Overall Length** 93"
- **Overall Width**
Siderails Up 39"
Siderails Down 37"
- **Lift Capacity**
Bed 500 lbs
Head Section 250 lbs
Foot Section 300 lbs
- **Height Range (to litter top)**
High 35.5" / Low 17.5"
- **Litter Positioning**
Backrest 0° - 70°
Pelvic Tilt 4°
Maximum Trend. 8°
- **Patient Surface** 33" x 81"
- **Mattress Thickness**
Head/Seat 5"
Foot 3"
- **Seat Depth**
Standard 12.5"
With Skoocher 12.5" - 8.5"
- **Caster Diameter**
Standard 6"